



Fil-Tech
CERTIFICATE OF ANALYSIS

Vendor Part No. QM1001

Part Description: Gold Coated 5 MHz Liquid Monitoring Crystals

It is hereby certified that all articles in the quantities as indicated above are in compliance with the following specifications:

Frequency: 4.950-5.050 MHz
Activity: Less than 15 Ohms measured at resonance
Contact Resistance: Less than 15 Ohms center to edge
Diameter: 1" \pm 0.001
Angle of Cut: 35 degrees 15' (AT) \pm 1'
Contour: Plano-Plano
Surface Roughness: Polished
Contact Composition: Chromium/Gold
Packaging: 100% Virgin PETG, Pharmaceutical Grade, recyclable plastic, Polystyrene base.

Signature: _____

Paula L. Becker
President/CEO
Fil-Tech, Inc.
6 Pinckney St.
Boston, MA 02114
617-227-1133 ph
617-742-0686 fax
paula@filtech.com
www.filtech.com