



Fil-Tech
CERTIFICATE OF COMPLIANCE

Customer Part No. QS3950

Part Description: Silver Coated 5 MHz Quartz Monitor Crystals.

Vendor Part No. QS3950

Quantity Inspected: pcs (100%)

It is hereby certified that all articles in the quantities as indicated above are in compliance with the following specifications.

Frequency: (series fundamental resonance) 4.990-5.002 MHz
Activity: Less than 15 Ohms measured at resonance
Contact Resistance: Less than 15 Ohms center to edge
Diameter: 0.490 +.000,-.002"
Angle of Cut: 35 degrees 15+/-3 minutes
Contour: Plano-Convex, 3+/-0.5 diopter curvature
Surface Roughness: Better than 10 microns rms
Contact Composition: 99.99+% pure Silver/Chromium
Packaging: 100% Virgin PETG, Pharmaceutical Grade,
recyclable plastic.

Signature: _____

Paula L. Becker, President/CEO

Fil-Tech, Inc.

6 Pinckney St.

Boston, MA 02114

617-227-1133 ph

617-742-0686 fax

paula@filtech.com

www.filtech.com