



**Fil-Tech**  
**CERTIFICATE OF COMPLIANCE**

**Customer Part No.** QS3954

**Part Description:** Stress Relieving Alloy Coated 5 MHz Quartz Monitor Crystals.

**Vendor Part No.** QS3954

**Quantity Inspected:** pcs (100%)

It is hereby certified that all articles in the quantities as indicated are in compliance with the following specifications.

**Frequency:** 4.990-5.002 MHz  
**Activity:** Less than 15 Ohms measured at resonance  
**Contact Resistance:** Less than 15 Ohms center to edge  
**Diameter:** 0.490+.000,-.002"  
**Angle of Cut:** 35 degrees 15+/-3 minutes  
**Contour:** Plano-Convex, 3+/-0.5 diopter curvature  
**Surface Roughness:** Better than 10 microns rms  
**Contact Composition:** Silver Aluminum Alloy/Chromium  
**Packaging:** 100% Virgin PETG, Pharmaceutical Grade,  
recyclable plastic.

Signature: \_\_\_\_\_

Paula L. Becker, President/CEO

Fil-Tech, Inc.

6 Pinckney St.

Boston, MA 02114

617-227-1133 ph

617-742-0686 fax

paula@filtech.com

www.filtech.com